

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

CB TEST CERTIFICATE

Product

Medical Switch Mode Power Supply

Name and address of the applicant

TDK-LAMBDA UK LTD
KINGSLEY AVE
ILFRACOMBE DEVON, EX34 8ES UNITED KINGDOM

Name and address of the manufacturer

TDK-LAMBDA UK LTD
KINGSLEY AVE
ILFRACOMBE DEVON, EX34 8ES UNITED KINGDOM

Name and address of the factory

Note: When more than one factory, please report on page 2

TDK-Lambda UK Ltd
Kingsley Avenue Ilfracombe Devon EX34 8ES
United Kingdom Additional Information on page 2

Ratings and principal characteristics

100-240Vac (Nominal), 45-63Hz, 3A (All AC models except as below)
88.9-240Vac (Nominal), 45-63Hz, 3A (Model NV1-1G000 only)
(see Model Differences in the report for details)

Trademark (if any)



Type of Customer's Testing Facility (CTF) Stage used

CTF Stage 3

Model / Type Ref.

NV-175 Series, NV175 Series
See Page 2

Additional information (if necessary may also be reported on page 2)

Risk Management was not included in this investigation

 Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1

As shown in the Test Report Ref. No. which forms part of this Certificate

E349607-D13-CB-2 issued on 2017-07-25

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2017-08-08

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

DK-65814-UL

Model Details:

NV-175 Series, NV175 Series See test report for details

Factories:

Panyu Trio Microtronic Co. Ltd
Shiji Industrial Estate Dongyong Nansha Guangzhou, Guangdong,
China

Additional Information:

Additionally evaluated to EN60601-1:2006/ A1:2013/ A12:2014 ; National Differences specified in the CB Test Report

Additional information (if necessary)



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2017-08-08

Signature:

Jan-Erik Storgaard