



Ref. Certif. No.

DK-121015-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Switch mode power supply
Name and address of the applicant	TDK-LAMBDA UK LTD KINGSLEY AVE ILFRACOMBE EX34 8ES UNITED KINGDOM
Name and address of the manufacturer	TDK-LAMBDA UK LTD KINGSLEY AVE ILFRACOMBE EX34 8ES UNITED KINGDOM
Name and address of the factory	TDK-LAMBDA UK LTD KINGSLEY AVE ILFRACOMBE EX34 8ES UNITED KINGDOM
Note: <i>When more than one factory, please report on page 2</i>	Additional Information on page 2
Ratings and principal characteristics	100-240Vac Additional Information on page 2
Trademark (if any)	TDK-LAMBDA
Customer's Testing Facility (CTF) Stage used	CTF Stage 3
Model / Type Ref.	NV175 Series, NV-175 Series, NV1-1G000
Additional information (if necessary may also be reported on page 2)	Additionally evaluated to: EN 61010-1:2010, EN 61010-1:2010/A1:2019 Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010, IEC 61010-1:2010/AMD1:2016
As shown in the Test Report Ref. No. which forms part of this Certificate	E331788-D1003-2/A0/C0-CB issued on 2021-11-16

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2021-11-18

Signature: Jan-Erik Storgaard

IEC**IECEE**

Ref. Certif. No.

DK-121015-UL**Factories:**

PANYU TRIO MICROTRONICS CO LTD
SHIJI INDUSTRIAL ESTATE
DONGYONG
NANSHA GUANGZHOU, GUANGDONG, 511453
CHINA

TRIO-TRONICS (THAILAND) LTD
7/295 MU. 6
MAP YANG PHON SUB-DISTRICT PLUAK DAENG DISTRICT, RAYONG PROVINCE,
THAILAND

Additional Ratings:

NV175 Series, NV-175 Series:
100-240Vac (Nominal), 90-264V (Full Tolerance), 45-440Hz, 3Arms
NV1-1G000 only:
88.9-240Vac (Nominal), 80-264V (Full Tolerance), 45-440Hz, 3Arms
(See test report Model Differences for details)

Additional information (if necessary)

- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2021-11-18

Signature:

Jan-Erik Storgaard