

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

Product
Produit

Name and address of the applicant
Nom et adresse du demandeur

Name and address of the manufacturer
Nom et adresse du fabricant

Name and address of the factory
Nom et adresse de l'usine

Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{eme} page

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Trademark (if any)
Marque de fabrique (si elle existe)

Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais
constructeur

Model / Type Ref.
Ref. De type

Additional information (if necessary may also be
reported on page 2)
Les informations complémentaires (si nécessaire,,
peuvent être indiqués sur la 2^{eme} page

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

As shown in the Test Report Ref. No. which forms part
of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue partie de ce Certificat

CERTIFICAT D'ESSAI OC

Medical Grade Power Supply

TDK-LAMBDA CORP
NAGAOKA TECHNICAL CENTER R&D DIV
2704-1 SETTAYA-MACHI
NAGAOKA-SHI
NIIGATA 940-1195 JAPAN

TDK-LAMBDA CORP
NAGAOKA TECHNICAL CENTER R&D DIV
2704-1 SETTAYA-MACHI
NAGAOKA-SHI
NIIGATA 940-1195 JAPAN

WUXI TDK-LAMBDA ELECTRONICS CO LTD
NO 6 XING CHUANG ER LU WUXI JIANGSU 214028
CHINA

Additional Information on page 2

See Page 2

HWS100A-xx/ME, HWS100A-xx/MEA, HWS100A-xx/MER,
HWS100A-xx/MERA, See Page 2

Additionally evaluated to EN 60601-1:2006 / A1:2013/ A12:2014;
National Differences specified in the CB Test Report

Additional Information on page 2

IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1

4787490305 issued on 2016-06-29

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2016-06-29

Original Issue Date: 2015-04-16

Signature:

Jolanta M. Wroblewska



Ref. Certif. No.

US-25008-A2-UL

Model Details:

HWS100A-xx/ME, HWS100A-xx/MEA, HWS100A-xx/MER, HWS100A-xx/MERA

Where xx represents the output voltage rating, and can be one of the following: 5, 12, 15, 24, or 48.

Factories:

TDK-LAMBDA MALAYSIA SDN BHD

PLO33 KAWASAN PERINDUSTRIAN SENAI 81400 SENAI
MALAYSIA

TDK-LAMBDA CORP

2704-1 SETTAYA-MACHI NAGAOKA-SHI NIIGATA-KEN 940-1195
JAPAN

ALPS LOGISTICS FACILITIES CO LTD

593-1 NISHI-OHASHI TSUKUBA-SHI IBARAKI-KEN 305-0831
JAPAN

SENDAN ELECTRONICS MFG CO LTD

1010 HABUSHIN NANTO-SHI TOYAMA-KEN 939-1756
JAPAN

TDK-LAMBDA MALAYSIA SDN BHD

LOT 2 & 3, BATU 9 3/4 KAWASAN PERINDUSTRIAN BANDAR BARU JAYA GADING 26070 KUANTAN
MALAYSIA

Ratings:

<Input>

HWS100A-xx/ME, HWS100A-xx/MEA, HWS100A-xx/MER, and HWS100A-xx/MERA : 100 - 240 V ac, 50 - 60 Hz, 1.4 A

<Output>

(1) HWS100A-5/ME, HWS100A-5/MEA, HWS100A-5/MER, and WS100A-5/MERA: 5 Vdc , 20 A

(2) HWS100A-12/ME, HWS100A-12/MEA, HWS100A-12/MER, and HWS100A-12/MERA: 12 Vdc , 8.5 A

(3) HWS100A-15/ME, HWS100A-15/MEA, HWS100A-15/MER, and HWS100A-15/MERA: 15 Vdc , 7.0 A

(4) HWS100A-24/ME, HWS100A-24/MEA, HWS100A-24/MER, and HWS100A-24/MERA: 24 Vdc , 4.5 A

(5) HWS100A-48/ME, HWS100A-48/MEA, HWS100A-48/MER, and HWS100A-48/MERA: 48 Vdc , 2.1 A

Additional Information:

The original report was modified to include the following changes/additions:

Add alternate components, update component table, see test report

The product has been evaluated with the exception of Risk Management assessment.

Additional information (if necessary)

Information complémentaire (si nécessaire)



UL (US), 333 Pflingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2016-06-29

Original Issue Date: 2015-04-16

Signature:

Jolanta M. Wroblewska