



Ref. Certif. No.

DK-88634-M1-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Switch mode power supply

Name and address of the applicant

TDK-LAMBDA UK LTD
Kingsley Avenue, Ilfracombe Devon, EX34 8ES
United Kingdom

Name and address of the manufacturer

TDK-LAMBDA UK LTD
Kingsley Avenue, Ilfracombe Devon, EX34 8ES
United Kingdom

Name and address of the factory

TDK-LAMBDA UK LTD
Kingsley Avenue, Ilfracombe Devon, EX34 8ES
United Kingdom

Note: When more than one factory, please report on page 2

Ratings and principal characteristics

100-240Vac nom, 45-63Hz, 4.9A rms max.
(see model differences for details of models and ratings)

[Additional Information on page 2](#)

Trademark / Brand (if any)



Type of Customer's Testing Facility (CTF) Stage used

CTF Stage 3

Model / Type Ref.

EFE300M Series
See Page 2

Additional information (if necessary may also be reported on page 2)

The report was revised to include technical modifications.
 [Additional Information on page 2](#)

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005

As shown in the Test Report Ref. No. which forms part of this Certificate

E349607-D1013-1/A1/C0-CB issued on 2020-03-31

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2020-04-09
Original Issue Date: 2019-10-17

Signature:

Jan-Erik Storgaard

Model Details:

EFE300M Series (see model differences for details of models and nomenclature)

Factories:

TRIO MICROTRONIC CO LTD
SHIJI INDUSTRIAL ESTATE, DONGYONG, PANYU, GUANGZHOU GUANGDONG
China

TDK-LAMBDA CORPORATION
2701 TOGAWA SETTAYA, NAGOAKA-SHI NIIGATA KEN, 940-1195
JAPAN

Additional Information:

The original report was modified to include following changes/additions:

- addition of EN standard to the report;
- update of country differences section.

The risk management requirements of the standard were not addressed. National Differences specified in the CB Test Report. Additionally evaluated to EN 60601-1:2006/A1:2013/A12:2014.

Additional information (if necessary)



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/names

Date: 2020-04-09

Original Issue Date: 2019-10-17

Signature:

Jan-Erik Storgaard