Ref. Certif. No.

SI-3328

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

CERTIFICAT D'ESSAI OC

Product Produit	Switching Power Supply for building-in
Name and address of the applicant Nom et adresse du demandeur	Arch Electronics Corp. 3F., No. 79, Sec. 1, Hsin Tai Wu Rd., Sijhih City, Taipei County 221, Taiwan
Name and address of the manufacturer Nom et adresse du fabricant	Arch Electronics Corp. 3F., No. 79, Sec. 1, Hsin Tai Wu Rd., Sijhih City, Taipei County 221, Taiwan
Name and address of the factory Nom et adresse de l'usine Note: *When more than one factory, please report on page 2* Note: *Lorsque il y plus d'une usine, veuillez utiliser la 2ème page*	Arch Electronics Corp. 3F., No. 79, Sec. 1, Hsin Tai Wu Rd., Sijhih City, Taipei County 221, Taiwan ☐ Additional Information on page 2
Ratings and principal characteristics Valeurs nominales et caractéristiques principales	I/P: 100-240 Vac; 47-63 Hz; 0,32-0,17 A O/P: ☒ Additional Information on page 2
Trademark (if any) Marque de fabrique (si elle existe)	TDK·Lambda
Type of Manufacturer's Testing Laboratories used Type de programme du laboratoire d'essais constructeur	/
Model / Type Ref. Ref. De type	KMx15-y ☒ Additional Information on page 2
Additional information (if necessary may also be reported on page 2) Les informations complémentaires (si nécessaire,, peuvent être indiqués sur la 2ème page	The risk management requirements of the standard were not addressed. ☐ Additional Information on page 2
A sample of the product was tested and found to be in conformity with Un échantillon de ce produit a été essayé et a été considéré conforme à la	IEC 60601-1:2005 (3rd Ed.)
As shown in the Test Report Ref. No. which forms part of this Certificate Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat	T223-0163/12, dated 2012-06-06

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

Slovenski institut za kakovost in meroslovje
Slovenian Institute of Quality and Metrology
Tržaška c. 2, SI-1000 Ljubljana, Slovenia
Product Certification Body is accredited by Slovenian Accreditation, Reg. No. CP-001

Date: 2012-06-06

Signature: Zoran Svetik

Model / Type Ref.:

KMx15-y

''x'' can be S, D or T

S= Single output

D= Dual output

T= Triple output

''y'' can be 3P3, 5, 7P35, 9, 12, 15, 24, 55, 1212, 1515, 51212 or 51515

Ratings and principal characteristics:

I/P: 100-240 Vac; 47-63 Hz; 0,32-0,17 A

O/P:

Model Name	Output Ratings (output dc voltage / output current)
KMS15-3P3	3,3 V / 3 A
KMS15-5	5 V / 3 A
KMS15-7P35	7,35 V / 2,04 A
KMS15-9	9 V / 1,666 A
KMS15-12	12 V / 1,25 A
KMS15-15	15 V / 1 A
KMS15-24	24 V / 0,625 A
KMD15-55	+ 5 V / 1,5 A; - 5 V / 1,5 A
KMD15-1212	+ 12 V / 0,625 A; - 12 V / 0,625 A
KMD15-1515	+ 15 V / 0,5 A; - 15 V / 0,5 A
KMT15-51212	+ 5 V / 2 A; + 12 V / 0,2 A; - 12 V / 0,2 A
KMT15-51515	+ 5 V / 2 A; + 15 V / 0,15 A; - 15 V / 0,15 A

Additional information (if necessary)
Information complémentaire (si nécessaire)

Date: 2012-06-06

Signature: Zoran Svetik