



Ref. Certif. No.

DK-45833-ULIEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC**CB TEST CERTIFICATE****CERTIFICAT D'ESSAI OC**Product
Produit

Switch Mode Power Supply

Name and address of the applicant
Nom et adresse du demandeurTDK-LAMBDA UK LTD
KINGSLEY AVE
ILFRACOMBE, EX34 8ES United KingdomName and address of the manufacturer
Nom et adresse du fabricantTDK-LAMBDA UK LTD
KINGSLEY AVE
ILFRACOMBE, EX34 8ES United KingdomName and address of the factory
Nom et adresse de l'usinePanyu Trio Microtronics Co Ltd
SHIJI INDUSTRIAL ESTATE
DONGYONG
NANSHA
GUANGZHOU, 511453 GUANGDONG ChinaNote: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{eme} page Additional Information on page 2
94.5 Vac to 240 Vac, (85-264Vac max. tolerance),
16A, 47-63 HzRatings and principal characteristics
Valeurs nominales et caractéristiques principales

TDK-Lambda

Trademark (if any)
Marque de fabrique (si elle existe)**TDK-Lambda**Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais
constructeur

SMT

Model / Type Ref.
Ref. De typeSeries: Alpha 800, Alpha 800W. Models: CA800 (followed by
various letters and numbers as defined in the model differences)
Series: Alpha 1000, Alpha 1000W., See Page 2Additional information (if necessary may also be
reported on page 2)
Les informations complémentaires (si nécessaire,,
peuvent être indiqués sur la 2^{eme} pageAdditionally evaluated to EN 60601-1:2006 / A1:2013; National
Differences specified in the CB Test Report Additional Information on page 2A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1

As shown in the Test Report Ref. No. which forms part
of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue partie de ce Certificat

E349607-D8-CB-1 issued on 2015-05-27

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2015-05-27

Signature:

Jan-Erik Storgaard



Ref. Certif. No.

DK-45833-UL

Model Details:

Models: CA1000 (followed by various letters and numbers as defined in the model differences), CA1250 12C_MF_PP 12F_PP 12F_PP 12F_PP, CA1250 12C_MF 12FF 12FF 12FF

Factories:

TDK-LAMBDA UK LTD
KINGSLEY AVE
ILFRACOMBE, EX34 8ES United Kingdom

Additional Information:

No Risk Management evaluation has been carried out on this product, therefore Essential Performance has not been specified, as this depends on the End Equipment

Additional information (if necessary)

Information complémentaire (si nécessaire)



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/nbcnames

Date: 2015-05-27

Signature:

Jan-Erik Storgaard