



Ref. Certif. No.

DK-70495-A1-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

CB TEST CERTIFICATE

Product

Medical Switch Mode Power Supply

Name and address of the applicant

TDK-LAMBDA UK LTD
KINGSLEY AVE ILFRACOMBE, DEVON EX34 8ES
UNITED KINGDOM

Name and address of the manufacturer

TDK-LAMBDA UK LTD
KINGSLEY AVE ILFRACOMBE, DEVON EX34 8ES
UNITED KINGDOM

Name and address of the factory

TDK-LAMBDA UK LTD
KINGSLEY AVE ILFRACOMBE, DEVON EX34 8ES
UNITED KINGDOM

Note: When more than one factory, please report on page 2

[Additional Information on page 2](#)

Ratings and principal characteristics

100-240Vac nom, 3A rms max, 45-63Hz.
(See Model Differences for details of ratings).

Trademark / Brand (if any)

TDK-Lambda

Type of Customer's Testing Facility (CTF) Stage used

CTF Stage 3

Model / Type Ref.

NVM-175, NVM175
See Page 2

Additional information (if necessary may also be reported on page 2)

The report was revised to include administrative modifications.
 [Additional Information on page 2](#)

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005

As shown in the Test Report Ref. No. which forms part of this Certificate

E349607-D1007-1/A1/C0-ULCB issued on 2020-03-24

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

Date: 2020-04-06
Original Issue Date: 2018-02-06

Signature:

Jan-Erik Storgaard

For full legal entity names see www.ul.com/ncbnames



Model Details:

NVM-175,NVM175 (see Model Differences for details of models and nomenclature)

Factories:

PANYU TRIO MICROTRONIC CO. LTD.

SHIJI INDUSTRIAL ESTATE DONGYONG, NANSHA, GUANGZHOU, GUANGDONG 511453
CHINA

Trio-Tronics (Thailand) Ltd.

6 Map Yang Phon, Sub-District Pluak Daeng District 7/295 Mu. , Rayong
THAILAND

Additional Information:

The original report was modified to include following changes/additions:

-addition of new factory.

National Differences specified in the CB Test Report.

The Risk Management requirements of the standard were not addressed.

Additional information (if necessary)



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/nbnames

Date: 2020-04-06

Original Issue Date: 2018-02-06

Signature:

Jan-Erik Storgaard