

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE**CERTIFICAT D'ESSAI OC**

Product
Produit

Switch mode power supply

Name and address of the applicant
Nom et adresse du demandeur

TDK-LAMBDA UK LTD
KINGSLEY AVENUE ILFRACOMBE DEVON EX34 8ES
UNITED KINGDOM

Name and address of the manufacturer
Nom et adresse du fabricant

TDK-LAMBDA UK LTD
KINGSLEY AVENUE ILFRACOMBE DEVON EX34 8ES
UNITED KINGDOM

Name and address of the factory
Nom et adresse de l'usine

TDK-LAMBDA UK LTD
KINGSLEY AVE ILFRACOMBE DEVON EX34 8ES
UNITED KINGDOM

Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{ème} page

Additional Information on page 2

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

NV175 Series; NV-175 Series:
100-240Vac (Nominal), 90-264V (Full Tolerance), 45-63Hz,
3Arms
NV1-1G000 only:
88.9-240Vac (Nominal), 80-264V (Full Tolerance), 45-63Hz,
3Arms

Trademark (if any)
Marque de fabrique (si elle existe)

TDK-Lambda

Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais
constructeur

SMT

Model / Type Ref.
Ref. De type

NV-175 Series, NV1-1G000, NV175 Series

Additional information (if necessary may also be
reported on page 2)
Les informations complémentaires (si nécessaire,,
peuvent être indiqués sur la 2^{ème} page

For further model and rating explanation see test report.
National Differences specified in the CB Test Report.

Additional Information on page 2

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

IEC 60601-1(ed.2), IEC 60601-1(ed.2);am1, IEC 60601-
1(ed.2);am2

As shown in the Test Report Ref. No. which forms
part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue partie de ce Certificat

E349607-A26-CB-1 issued on 2014-12-05

This CB Test Certificate is issued by the National Certification Body

Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

Date: 2014-12-08

Original Issue Date: 2012-06-13

Signature:

Jan-Erik Storgaard

For full legal entity names see www.ul.com/ncbnames



Ref. Certif. No.

DK-26602-A2-UL

Factories:

PANYU TRIO MICROTRONIC CO. LTD.
SHIJI INDUSTRIAL ESTATEDONGYONG NANSHA
CHINA

Additional Information:

The original report was modified to include the following changes/additions:

1. Addition/deletion of multilayer PWBs to critical component list.
2. Critical component certificate updates.
3. Change factories.
4. Correction/Addition to CCL components.
5. Addition of 18V channel 1 with fan output. (thermal test carried out)

Additional information (if necessary)

Information complémentaire (si nécessaire)



UL (US), 333 Pflingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2014-12-08

Original Issue Date: 2012-06-13

Signature:

Jan-Erik Storgaard