



Ref. Certif. No.

DK-64715-A2-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Switch Mode Power Supply
Name and address of the applicant	TDK-LAMBDA UK LTD KINGSLEY AVE ILFRACOMBE DEVON EX34 8ES UNITED KINGDOM
Name and address of the manufacturer	TDK-LAMBDA UK LTD KINGSLEY AVE ILFRACOMBE DEVON EX34 8ES UNITED KINGDOM
Name and address of the factory	TDK-LAMBDA UK LTD KINGSLEY AVE ILFRACOMBE DEVON EX34 8ES UNITED KINGDOM
Note: <i>When more than one factory, please report on page 2</i>	<input checked="" type="checkbox"/> Additional Information on page 2
Ratings and principal characteristics	XMS350, XMS-350: 100-240Vac nom, 47-63Hz, 5.3A rms max. XMS500, XMS-500: 100-240Vac nom, 47-63Hz, 7A rms max. (see test report for details)
Trademark (if any)	
Type of Customer's Testing Facility (CTF) Stage used	CTF Stage 3
Model / Type Ref.	XMS350 or XMS-350 and XMS500 or XMS-500 series, See Page 2
Additional information (if necessary may also be reported on page 2)	<input checked="" type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1(ed.2), IEC 60601-1(ed.2);am1, IEC 60601-1(ed.2);am2
As shown in the Test Report Ref. No. which forms part of this Certificate	E349607-A41-CB-2 issued on 2018-06-15

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

Date: 2018-06-20
Original Issue Date: 2017-07-06

Signature:

Jan-Erik Storgaard

For full legal entity names see www.ul.com/ncbnames



Ref. Certif. No.

DK-64715-A2-UL

Model Details:

XMS350 or XMS-350 and XMS500 or XMS-500 series switch mode power supplies
(see report Model Differences for details of nomenclature)

Factories:

PANYU TRIO MICROTRONIC CO. LTD
SHIJI INDUSTRIAL ESTATE DONGYONG NANSHA GUANGZHOU GUANGDONG
CHINA

Additional Information:

National Differences specified in the CB Test Report.

The original report was modified to include the following changes/additions:

- Addition of alternates to the CCL.
- Corrections to CCL.
- Testing of alternate discharge resistors.
- Enclosures updated where applicable.

Additional information (if necessary)



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2018-06-20
Original Issue Date: 2017-07-06

Signature:

Jan-Erik Storgaard